



DEXTERITY



LIGHTLY
POWDERED



CHEMICAL
RESISTANT

LATEX THIN 44-140



CHARACTERISTICS

- Lightly powdered latex glove
- Smooth inside
- Supple and pleasant to wear
- Very easy to put on thanks to rolled edge
- Suitable for the right and left hand (ambidextrous)
- Suitable for contact with many foods
- Meets the EN 455 requirement for medical examination gloves
- Length: 240 mm and thickness: 0.08 mm
- Weight: 5.8 grams (+/- 0.3 grams)
- AQL 1.5 (EN 374)
- Gloves are for single use only
- Packed per 100 pieces in a dispenser

Article number: 1.44.140.00

SUITABLE FOR ACTIVITIES IN E.G.

- Food industry
- Agri
- Hospitality
- Maintenance
- Cleaning
- Automotive

COLOUR

Natural

SIZES

7/S to 10/XL incl.

PACKAGING

- 100 pieces per dispenser
- 10 dispensers in a case

EN ISO 374-1:2016
+A1:2018/Type C



EN ISO 374-5:
2016




EN 455

CE 1639
EN 420:2003+A1:2009

PRODUCT INFORMATION

SIZE	ARTICLE NO.	EAN CODE 100 PIECES (DISPENSER)	EAN CODE 10 DISPENSERS (CASE)
7/S	1.44.140.07	8718249008085	8718249008092
8/M	1.44.140.08	8718249008108	8718249008115
9/L	1.44.140.09	8718249008122	8718249008139
10/XL	1.44.140.10	8718249008146	8718249008153



CLARIFICATION OF PICTOGRAMS

Protection against chemicals and micro-organisms

EN ISO 374-1:2016/Type B		
EN ISO 374-1:2016 /Type A  K	Chemicals 40% Sodium Hydroxide (K)	EN 16523:2015 permeation level 6
EN ISO 374-4:2019		
CAS number (K) 1310-73-2	Class Inorganic base	
Chemical 40% Sodium Hydroxide (K)	Average decrease % -33.0%	

Resistance to bacteria & fungi and virus

EN ISO 374-5:2016		
EN ISO 374-2:2014  VIRUS	Resistance to bacteria & fungi PASS	Virus resistance PASS
EN ISO 374-2:2014		
Air leak test - Pass	Water leak test - Pass	

EN ISO 374-1:2016+A1:2018 Permeation levels are based on the following lead times:

Performance level	1	2	3	4	5	6
Minimum breakthrough times (in min.)	>10	>30	>60	>120	>240	>480

STORAGE REQUIREMENTS

The gloves should be stored in a clean, cool and dry place and not kept compressed in the original packaging. Do not expose the gloves to direct sunlight. Make sure that the packaging and the gloves are not damaged when shipped.

TESTING INSTITUTE

These gloves are certified by: SATRA Technology Europe Ltd (Notified Body no. 2777), Bracetown Business Park, Cloness, Dublin D15 YN2P, Ireland.

DECLARATION OF CONFORMITY

For a copy of the declaration of conformity, please refer to the following web link: www.oxxa.work/doc

RELATED ITEMS



LATEX-STRONG 44-160
ART. NO. 1.44.160.00



LATEX-STRONG 44-161
ART. NO. 1.44.161.00

LATEX-STRONG 44-215 - ART. NO. 1.44.215.00

YOUR SUPPLIER: