



FIT



FEEL



CHEMICAL
RESISTANT

E-FUTURE- NITRILE 52-500



CHARACTERISTICS

- Unpowdered non-sterile disposable nitrile glove
- Skin-friendly gloves and pure formula: no chlorine
- no rubber latex proteins and no rubber accelerators (accelerators)
- These gloves feature embossed fingertips for extra grip in wet or dry applications
- Smooth interior and rolled cuff ensure easy donning and doffing of the glove
- Suitable for the right and left hand (ambidextrous)
- The comfortable and strong material ensures a pleasant wearing experience, even over long periods of time
- Silicone-free
- Length: 240 mm
- Thickness: cuff 0.05 mm / palm 0.06 mm / fingers 0.07 mm
- Weight: 3.5 g (+/- 0.2 g)
- AQL 1.5 (EN 374)
- Suitable for contact with many foodstuffs
- Meets requirement EN 455 for medical examination gloves
- Packed per 100 pieces in a dispenser

Article number: 1.52.500.00

SUITABLE FOR ACTIVITIES IN E.G.

- Food industry
- Agriculture
- Automotive
- Hospitality
- Chemical industry

COLOR

Green

SIZES

7/S to 10/XXL

PACKAGING

- 100 pieces per dispenser
- 10 dispensers per outerbox

EN ISO 374-1:2016
+A1:2018/Type B



EN ISO 374-5:
2016



EN 455

CE 2777
EN ISO 21420:2020

PRODUCT INFORMATION

SIZE	ARTICLE NO.	EAN CODE 100 PIECES (DISPENSER)	EAN CODE 10 DISPENSER (OUTER BOX)
7/S	1.52.500.07	8718249072192	8718249072239
8/M	1.52.500.08	8718249072208	8718249072246
9/L	1.52.500.09	8718249072215	8718249072253
10/XL	1.52.500.10	8718249072222	8718249072260

CLARIFICATION OF PICTOGRAMS

Protection against chemicals and micro-organisms

EN ISO 374-1:2016/Type B		
EN ISO 374-1:2016 /Type A	Chemicals	EN 16523:2015 permeation level
	40% Sodium Hydroxide (K) 30% Hydrogen peroxide (P) 37% Formaldehyde (T)	6 4 6
EN ISO 374-4:2019		
Cas number (K) 1310-73-2 (P) 7722-84-1 (T) 50-00-0	Class Inorganic base Peroxide Aldehyde	
Chemicals	Average degredation %	
40% Sodium Hydroxide (K) 30% Hydrogen peroxide (P) 37% Formaldehyde (T)	-3.9% 8.1% 2.9%	

Resistance to bacteria, fungi and viruses

EN ISO 374-5:2016		
EN ISO 374-2:2014	Resistance to bacteria & fungi	Resistance to Virus
	PASS	PASS
VIRUS		

EN ISO 374-2:2014	
Air leak test - Pass	Water leak test - Pass

EN ISO 374-1:2016 Permeation level are based on the following lead times:

Performance level	1	2	3	4	5	6
Minimum transit times (in min.)	>10	>30	>60	>120	>240	>480

STORAGE CONDITIONS

The gloves should be kept in a clean, cool and dry place and not kept compressed in their original packaging. Do not expose the gloves to direct sunlight. Make sure that the packaging and the gloves are not damaged during shipping.

TESTING INSTITUTE

These gloves have been certified by: SATRA Technology Europe Ltd (Notified Body no. 2777), Bracetown Business Park, Clonee, Dublin D15 YN2P, Ireland.

DECLARATION OF CONFORMITY

For a copy of the declaration of conformity, we refer you to the following link: www.oxxa.work/doc

YOUR SUPPLIER: